


AVOID ACCUMULATION OF DIRT IN THE PRESSURE CHANNEL CAUSED
BY VERTICAL INSTALLATION (CONNECTOR FACING DOWNWARD)
VERSCHMUTZUNG DES DRUCKKANALES BEI VERTIKALEM EINBAU
(STECKER NACH UNTEN) VERMEIDEN
EVITER L'INTRUSION DE POUSSIÈRES DANS LE CANAL SOUS PRESSION
(CONNECTEUR VERS LE BAS)
EVITA L'ACCUMULAZIONE DELLO SPORCO NEL CANALE DELLA PRESSIONE
CAUSATO DALL'INSTALLAZIONE VERTICALE (CONNETTORE VERSO IL BASSO)

COMPLETE DEVICE WITH MOUNTED COUPLING
KOMPLETTES GERAET MIT MONTIERTER KUPPLUNG *****
APPAREIL COMPLET AVEC L' ACCOUPLEMENT MONTE
APPARECCHIO COMPLETO CON ACCOPPIAMENTO MONTATO

PLUG-IN CONNECTION
 ***** STECKVERBINDUNG (EL.) DIN 72 585-A1-3.1-Sn/K2
 CONNEXION A FICHE
 ALLACCIAMENTO A SPINA

HYSTERESIS
HYSTERESE
HYSTERESIS ≤ 0.2%FS
ISTERESI

F